



Office of the Vice President
and Chief Financial Officer
Procurement Services
<https://procurement.ufl.edu/>

971 Elmore Drive
PO Box 115250
Gainesville, FL 32611-5250
(352) 392-1331 Fax 352-392-8837

February 10, 2023

ADDENDUM#1 TO INVITATION TO NEGOTIATE ITN23AW-123

TITLE: Radiography Systems and Maintenance

Bid Opening: February 22, 2023, at 3:00 p.m. EST

This addendum shall be considered part of the Contract Documents for the above mentioned **ITN23AW-123** as though it had been issued at the same time and incorporated integrally therewith. Where provisions of the following supplementary data differ from those of the original document, this addendum shall govern and take precedence. All other terms, conditions, and regulations will apply.

This addendum consists of:

- Responses to Supplier questions and requests for clarification received prior to February 2nd, 2023, at 5:00 p.m.

Ann Wright
Procurement Agent II

PLEASE ACKNOWLEDGE RECEIPT OF THIS ADDENDUM 1 AND RETURN WITH YOUR BID. FAILURE TO ACKNOWLEDGE THIS ADDENDUM COULD CONSTITUTE REJECTION OF YOUR BID.

Vendor Name

Vendor Address

Signature

Supplier Questions and Answers

1.2 Scope of Work

Q1. How many systems are needed?

A1. As per section 1.2 Scope of Work, there is a current need for three systems, however other units may be purchased in the future, either singly or in multiples.

Q2. Does the bid need to add the cost of the de-installation of the old Bucky cabinets as well as the installation of the new ones?

A2. Yes. Please include this in Tab 2 Pricing Proposal.

Q3. Does the snap on detector covers need to be included in the price?

A3. Yes, please see section 1.2 Scope of Work, Specifications.

Q4. The scope of work is asking for a Bucky Tray. Is this for an existing X-ray table? If so what brand and model is it for?

A4. Yes, the bucky trays are for the existing X-ray tables. The current bucky trays hold the 50g wired Canon plates. The new Canon wireless plates would need a new bucky tray to accommodate the smaller wireless plate.